

Vancouver Coastal Health

1200 - 601 W. Broadway Vancouver, BC V5Z 4C2

Phone Number: (604) 675-3800 Fax Number: (604) 736-8651

ENVIRONMENTAL HEALTH INSPECTION REPORT

Facility Information: Fairview Electrolysis Clinic Evelyn Geatros 314 - 1037 Broadway W Vancouver, BC V6H 1E3 Site Phone: 604-730-2624	Facility Number: 43065 Report Number: 9 Inspection Date: June 27, 2012 Community: 23
Facility Category: Personal Services, Skin Care Salon Inspection Type: Routine: N/A Action(s) Taken: Information Exchanged Delivery Method: Electronic Mail	
Opening Comments and Observations: Met with Evelyn Geatros, Operator Inspection conducted with Debby Peng, student EHO Services provided: Electrolysis only	

Observations and Recommendations

Personal Service Establishments: In Compliance - Personal Services

1 New: 0 / Resolved: 0

Bead sterilizer is used for sterilizing forceps/tweezers - using 230 C for 10 minutes - sterilizer is equipped with a thermometer- a separate timer is used to monitor for the minimum 10 minutes

Hand washing sink, fully equipped with liquid soap and paper towels in dispensers is available for use, in the room

Bleach solution in bowl- >1000 PPM chlorine

Public washrooms- in the hallway, common to other units on the floor.

Needles used for electrolysis- are sterile and single use only- used needles are discarded in a rigid plastic container.

Sanitation of premises is good.

Maintenance of premises is satisfactory.

Hand washing facilities are satisfactory.

Cleaning and disinfection of common-use tools and instruments appear to be satisfactory.

Closing Comments:

Evelyn has provided her disinfection/sterilization procedures, as requested.

- Needles are single use only - and NOT sterilized and re-used.

I have read and understood this report _____	Inspector _____
	3022 Harry Dhaliwal, Environmental Health Officer

Compliance Legend: IC = In Compliance, CA = Compliance Achieved During Inspection, NIC = Not In Compliance, NA = N/A